

Westlake weekly air monitoring update (Friday, August 15, 2014).

Goal: Develop an air monitoring plan to sample for landfill constituents (COC's, RIM), in conjunction with the PRP, prior to and during the construction of the isolation barrier (IB) being built to contain the RIM in OU-1, that could be impacted by the SSE.

Admin: TO/IGCE amendments have been submitted for the next 90 day monitoring period.

Data: The AreaRaes at the St. Charles location showed "pump failure" on the sensors. This was due to the filters being clogged. All AreaRaes at all stations have been cleaned and changed out. The plan is to change the filters bi-monthly.

QAPP: Latest changes are addressed in the DQO-Draft07282014.docx.

Equipment: Two Multi-Raes will be available at the trailer once the barrier construction starts.

ECAM: Still waiting on NRC license to possess/use radioactive materials.

Site File: The OSC.gov website was updated with current weekly reports, site photos and commentary.

Issues: FOIA is still in a "hold" status.

Media: No media interest generated at the air monitoring site this week.

Westlake Air Monitoring Update.08152014.docx